



Vodovodna 95
 1000 Ljubljana
 SLOVENIJA

T: +386 (0)1 428 44 93
 E: nfl@policija.si
www.policija.si


ANALYTICAL REPORT

3,4-Methylenedioxy-PV9 (C₁₉H₂₇NO₃)

1-(2H-1,3-benzodioxol-5-yl)-2-(pyrrolidin-1-yl)octan-1-one

Remark – other active cpd. detected: **none**

Sample ID:	1716-16
Sample description:	powder - white-off
Sample type:	RM-reference material
Comments ¹ :	CAY Lot#047063312; RESPONSE -purchasing
Date of entry:	1/6/2017

Substance identified-structure ² (base form)	
Systematic name:	1-(2H-1,3-benzodioxol-5-yl)-2-(pyrrolidin-1-yl)octan-1-one
Other names:	MDPV three carbon homolog; Methylenedioxy-Pyrovalerone three carbon homolog
Formula (per base form)	C ₁₉ H ₂₇ NO ₃
M _w (g/mol)	317,43
Salt form:	HCl
StdInChIKey (for base form)	SUZRHPREJMNOG-UHFFFAOYSA-N
Other active cpd. detected	none
Add.info (purity..)	98%

¹ This report has been produced with the financial support of the Prevention of and fight against crime Programme of the European Union (grant agreement number JUST/2013/ISEC/DRUGS/AG/6413). The contents of this report are the sole responsibility of the National Forensic Laboratory and can in no way be taken to reflect the views of the European Commission.

² Created by OPSIN free tool: <http://opsin.ch.cam.ac.uk/> DOI: 10.1021/ci100384d



Report updates

date	comments (explanation)

Supporting information

Analytical technique:	applied	remarks
GC-MS (EI ionization)	+	NFL GC-RT (min): 9,31 BP(1): 168; BP(2): 169, BP(3) :149,
FTIR-ATR	+	direct measurement
GC-IR (condensed phase)	+	always as base form

1. GC-MS (Agilent): GC-method is RT locked to tetracosane (9.258 min). Injection volume 1 ml and split mode (1:50). Injector temperature: 280 °C. Chromatographic separation: on column HP1-MS (100% dimethylpolysiloxane), length 30 m, internal diameter 0.25 mm, film thickness 0.25 µm. Carrier gas He: flow-rate 1.2 ml/min. GC oven program: 170 °C for 1 min, followed by heating up to 190 °C at rate 8 °C/min, then heating up to 293 °C at a rate of 18 °C/min, hold for 7.1 min, then heating at 50 °C/min up to 325 °C and finally 6.1 min isothermal. MSD source EI = 70 eV. GC-MS transfer line T= 235 °C, source and quadropole temperatures 280 °C and 180 °C, respectively. Scan range m/z scan range: from 50 (30 until 6 min.) to 550 (300 until 6 min) amu.

2. FTIR-ATR (Perkin Elmer): scan range 4000-400 cm⁻¹; resolution 4cm⁻¹

3. GC- (MS)-IR condensed phase (GC-MS (Agilent) & IR (Spectra analyses-Danny))

GC-method: Injection volume 1 ml and split mode (1:5). Injector temperature 280 °C. Chromatographic separation as above (1). Split MS : IR = 1 : 9.

MSD source EI = 70 eV. GC-MS transfer line T= 235 °C, source and quadropole temperatures 280 °C and 180 °C, respectively. Scan range m/z scan range: from 50 (30 until 6 min.) to 550 (300) amu.


IR (condensed (solid) phase): IR scan range 4000 to 650, resolution 4 cm⁻¹.

4. HPLC-TOF for exact monoisotopic mass and empirical formula control - results are not shown in the report.


FIGURES OF SPECTRA

MS (EI)

Abundance



FTIR-ATR (sample as received)



IR-Condensed phase

