


ANALYTICAL REPORT

3F-PCP (C₁₇H₂₄FN)

1-[1-(3-fluorophenyl)cyclohexyl]piperidine

Remark – other NPS detected: none

Sample ID:	2186-20
Sample description:	powder
Sample type:	test purchase /ISF projekt (NFL-SI)
Date of entry (DD/MM/YYYY) into NFL database:	20/10/2020
Report updates (if any) will be published here:	http://www.policija.si/apps/nfl_response_web/seznam.php

Substance identified - structure ¹ (base form)	
Systematic name	1-[1-(3-fluorophenyl)cyclohexyl]piperidine
Other names	3-Fluoro-PCP
Formula (per base form)	C ₁₇ H ₂₄ FN
M _w (g/mol)	261,38
Salt form/anions detected	HCl
StdInChIKey (per base form)	PFPLGKFWWBXTNP-UHFFFAOYSA-N
Other NPS detected	none
Additional info (purity..)	>98% purity of a sample based on ¹ H NMR

¹ Created by OPSIN free tool: <http://opsin.ch.cam.ac.uk/> DOI: 10.1021/ci100384d

Report updates

date	comments (explanation)

Instrumental methods (if applied) in NFL

1. GC-MS (Agilent): GC-method is RT locked to tetracosane (9.258 min). Injection volume 1 µl and split mode (1:50). Injector temperature: 280 °C. Chromatographic separation: on column HP1-MS (100% dimethylpolysiloxane), length 30 m, internal diameter 0.25 mm, film thickness 0.25 µm. Carrier gas He: flow-rate 1.2 ml/min. GC oven program: 170 °C for 1 min, followed by heating up to 190 °C at rate 8 °C/min, then heating up to 293 °C at a rate of 18 °C/min, hold for 7.1 min, then heating at 50 °C/min up to 325 °C and finally 6.1 min isothermal. MSD source EI = 70 eV. GC-MS transfer line T= 235°C, source and quadropole temperatures 280°C and 180°C, respectively. Scan range m/z scan range: from 50 (30 until 6 min.) to 550 (300 until 6 min) amu.

2. HPLC-TOF (Agilent): 6230B TOF with Agilent 1260 Infinity HPLC with binary pump, column: Zorbax Eclipse XDB-C18, 50 x 4.6 mm, 1.8 micron. Mobile phases (A) 0.1% formic acid and 1mM ammonium formate in water; (B) 0.1% formic acid in methanol (B). Gradient: starting at 5% B, changing to 40% B over 4 min, then to 70% over 2 min and in 5 min to 100%, hold 1 min and back to 5%, equilibration for 1.7 min. The flow rate: 1.0 ml/min; Injection volume 1 µl. MS parameters: 2GHz, Extended Dynamic range mode to a maximum of 1700 amu, acquisition rate 1.30 spectra/sec. Sample ionisation: by Agilent Jet Stream technology (Dual AJS ESI). Ion source: positive ion scan mode with mass scanning from 82 to 1000 amu. Other TOF parameters: drying gas (N₂) and sheath temperature 325 °C; drying gas flow rate 6 l/min; sheath gas flow rate 8 l/min; nebulizer 25 psig; Vcap. 4000 V; nozzle 2000 V; skimmer 65 V; fragmentor 175 V and Octopole RF 750 V.

3.FTIR-ATR (Perkin Elmer): scan range 4000-400 cm⁻¹; resolution 4cm⁻¹

4. GC- (MS)-IR condensed phase (GC-MS (Agilent) & IR (Spectra analyses-Danny)

GC-method: Injection volume 1 µl and split mode (1:5). Injector temperature 280 °C. Chromatographic separation as above **(1)**. Split MS : IR = 1: 9.

MSD source EI = 70 eV. GC-MS transfer line T= 235°C, source and quadropole temperatures 280°C and 180°C, respectively. Scan range m/z scan range: from 50 (30 until 6 min.) to 550 (300) amu.

IR (condensed (solid) phase): IR scan range 4000 to 650, resolution 4 cm⁻¹.

5. IC (anions) (Thermo Scientific, Dionex ICS 2100), Column: IonPac AS19, 2 x 250mm; Eluent: 10mM KOH from 0 to 10 min, 10-58 mM from 10 to 40min; Flow rate: 0.25 ml/min; Temperature: 30°C; Suppressor: AERS 500 2mm, suppressor current 13mA; Inj. Volume: 25 µl

Supporting information


Solubility in	result/remark
CH ₂ Cl ₂	soluble
MeOH	soluble
H ₂ O	partially

Analytical technique:	applied	remarks
GC-MS (EI ionization)	+	NFL GC-RT (min): 6,61 BP(1): 218; BP(2): 109,BP(3) :204,
HPLC-TOF	+	Exact mass (theoretical): 261,1893; measured value Δppm:-0,77; formula:C17H24FN
FTIR-ATR	+	direct measurement (sample as received)
FTIR (solid phase) always as base form	+	
IC (anions)	+	
NMR (in FKKT)	+	
validation		
other		


ANALYTICAL RESULTS

MS (EI)


Abundance



FTIR-ATR - direct measurement (sample as received)



IR (solid phase – after chromatographic separation)



TOF REPORT


Data File	3F-PCP_2186-20.d	Sample Name	ID-2186-20
Sample Type	Sample	Position	P2-A3
Instrument Name	6230B TOF LC-MS	User Name	
Acq Method	general-15_01_2020-XDB-C18-ESI+.m	Acquired Time	9/2/2020 7:57:15 PM
IRM Calibration Status	Success	DA Method	a-Drugs_NFL.m
Comment			

Compound Table


Label	Compound Name	MFG Formula	Obs. RT	Obs. Mass
Cpd 3: 3F-PCP	3F-PCP	C17 H24 F N	6.1	261.1895

Name	Obs. m/z	Obs. RT	Obs. Mass	DB RT	DB Formula	DB Mass	DB Mass Error (ppm)
3F-PCP	262.1967	6.1	261.1895	6.1	C17 H24 F N	261.1893	-0.77


Compound Chromatograms



MFE MS Zoomed Spectrum



MS Zoomed Spectrum



MS Spectrum Peak List


Obs. m/z	Charge	Abund	Formula	Ion/Isotope
262.1967	1	28772172	C17 H24 F N	(M+H)+
263.2001	1	5518122.06	C17 H24 F N	(M+H)+

--- End Of Report ---

Peak Integration Report

Sample Name:	2186_IC	Inj. Vol.:	25,00
Injection Type:	Unknown	Dilution Factor:	1,0000
Program:	ANIONI	Operator:	kemija
Inj. Date / Time:	02-sep-2020 / 12:11	Run Time:	43,00

No.	Time min	Peak Name	Peak Type	Area $\mu\text{S}\cdot\text{min}$	Height μS	Amount mg/L
1	8,77	Chloride	BMB	25,877	122,042	n.a.
TOTAL:				25,88	122,04	0,0




University
of Ljubljana


Faculty of Chemistry
and Chemical Technology



R E P O R T

Contract No.	C1714-19-460155 (Republic of Slovenia, Ministry of the Interior, POLICE)
Sample ID:	2186-20
Received date:	September 10, 2020
Our notebook code:	NFL-2186-20
NMR sample preparation:	16.7 mg dissolved in 0.7 mL DMSO- <i>d</i> ₆
NMR experiments:	¹ H, ¹³ C, ¹ H- ¹ H <i>gs</i> -COSY, ¹ H- ¹³ C <i>gs</i> -HSQC, ¹ H- ¹³ C <i>gs</i> -HMBC, ¹ H- ¹⁵ N <i>gs</i> -HMBC, ¹⁹ F
Proposed structure with formula, exact mass, molecular weight:	 Chemical Formula: C₁₇H₂₅FN⁺ Exact Mass: 262,1966 Molecular Weight: 262,3919
Chemical name:	<i>N</i> -protonated 1-(1-(3-fluorophenyl)cyclohexyl)piperidine
Comments:	- Structure elucidation based on 1D and 2D NMR spectra and HRMS. - >98% purity of a sample based on ¹H NMR spectrum.
Supporting information:	Copies of ¹ H and ¹³ C NMR spectra, ¹ H and ¹³ C FIDs.
Principal investigator:	Prof. Dr. Janez Košmrlj
Date of report:	October 15, 2020

NFL-2186-20
1H



```


Current Data Parameters
NAME      NFL-2186-20
EXPNO     1
PROCNO    1

F2 - Acquisition Parameters
Date_     20200925
Time      15.50
INSTRUM   spect
PROBHD    5 mm PABBO BB-
PULPROG   zg30
TD         65536
SOLVENT    DMSO
NS         32
DS         2
SWH         10000.000 Hz
FIDRES     0.152588 Hz
AQ         3.276799 sec
RG         80.6
DW         50.000 usec
DE         6.50 usec
TE         296.0 K
D1         1.00000000 sec
TD0        1

===== CHANNEL f1 =====
SFO1       500.130085 MHz
NUC1       1H
P1         8.70 usec
PL1        26.00000000 W

F2 - Processing parameters
SI         65536
SF         500.1300047 MHz
WDW        EM
SSB        0
LB         0.30 Hz
GB         0
PC         1.00
  
```

NFL-2186-20
13C



Current Data Parameters

NAME	NFL-2186-20
EXPNO	2
PROCNO	1

F2 - Acquisition Parameters

Date_	20200925
Time	22.50
INSTRUM	spect
PROBHD	5 mm PABBO BB-
PULPROG	zgpg30
TD	65536
SOLVENT	DMSO
NS	4096
DS	4
SWH	29761.904 Hz
FIDRES	0.454131 Hz
AQ	1.1010048 sec
RG	2050
DW	16.800 usec
DE	6.50 usec
TE	296.0 K
D1	1.00000000 sec
D11	0.03000000 sec
TD0	1

===== CHANNEL f1 =====

SFO1	125.7703637 MHz
NUC1	13C
P1	8.70 usec
PLM1	122.0000000 W

===== CHANNEL f2 =====

SFO2	500.1320005 MHz
NUC2	1H
CPDPRG12	waltz16
PCPD2	80.00 usec
PLW2	26.00000000 W
PLM12	0.30046001 W
PLM13	0.15113001 W

F2 - Processing parameters

SI	32768
SF	125.7578446 MHz
WDW	EM
SSB	0
LB	1.00 Hz
GB	0
PC	1.40