




Analytical data for alpha-PPP



1-phenyl-2-(pyrrolidin-1-yl)propan-1-one

GC-MS

Rt.: 7.43 min



An Agilent 6890N Network GC system set up with Agilent HP-5MS (length: 30 m, diameter: 0.25 mm, film: 0.25 mm) coupled to an Agilent 5973 Network Mass Selective Detector (scan range m/z 35 – m/z 500) was used. Samples were subjected to electron ionization (EI) mode. GC-MS conditions: HP-5MS column was temperature programmed from 100 °C (which was held for 2 minutes) to 280 °C at 20 °C/min, 280 °C was held for 3 minutes, then to 315 °C at 25 °C/min, the temperature was stated at 315 °C for 12 minutes. The carrier gas was helium. Tribenzyl-amine was applied as an internal standard (locked to 10.8 minutes). Data handling was carried out with GC/MSD ChemStation software.

ATR-FT-IR (powder)

